

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Disinfectant						
		Cidex	Cidex OPA	Gigasept AF	Gigasept FF	Virkon	Cidex Plus	Wavicide-01
Manufacturer		Johnson and Johnson	Johnson and Johnson	Schülke&Mayr UK Ltd.	Schülke&Mayr UK Ltd.	Dupont (Antec)	Johnson and Johnson	Medical Chemical Corp.
Level		HLD	HLD	-	HLD	HLD	HLD	HLD
Active Ingredients		Glutaraldehyde	Orthophthalaldehyde	Succinc dialdehyde	Succinc dialdehyde	Potassium	Glutaraldehyde	Glutaraldehyde
FDA Approval		Y	Y	Y	Y	Y	Y	Y
Disinfectant Type		Liquid	Liquid	Liquid	Liquid	Powder	Liquid	Liquid
Convex	C1-6/C1-6T/C1-6i/SC1-6/SC1-4HS/SC1-4H/SC1-6H/ SC2-9H/SC1-7H	Δ	O	Δ	X	O	O	O
	C1-6CT	Δ	O	Δ	X	Δ	Δ	O
Micro-Convex	C5-8	O	O	O	O	Δ	O	O
	C5-8N/C5-8NT/SC2-11H	Δ	O	Δ	X	O	O	O
Linear	L3-8/L3-8i/L3-12/L3-12T/L3-12i/L3-12H/L3-12H ^{WD} /L3-2X/L8-17/L8-17i/L8-17X/L3-15H	Δ	O	Δ	X	O	O	O
	L10-25H/SL3-19H/SL3-19X	Δ	O	Δ	O	O	O	Δ
Phased	MP1-5X/ SP1-5/SP1-5i/SP1-5X/P1-5CT/SP1-5T/SP3-8/SP3-8T	Δ	O	Δ	X	O	O	O
Volume	VC1-6/VC1-6T/SVC1-6/SVC1-6H/VE3-10H/SVC1-8H	Δ	O	Δ	O	O	O	Δ
Endo	EV3-10/EV3-10T/EV3-10H/EV3-10X	Δ	O	Δ	X	O	O	O
	EC3-10/EC3-10T/EC3-10H/EC3-10X/EC(V)2-11H	Δ	O	Δ	X	O	O	O
	EN3-10	O	O	Δ	Δ	Δ	O	O
	E3-10	Δ	O	Δ	Δ	O	O	Δ
	E3-10H	Δ	O	Δ	Δ	O	O	Δ
IO	IO3-12/IO8-17/IO8-17T	Δ	O	Δ	X	O	O	O
CW	CW2.0/CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	Δ	O	Δ	Δ	Δ	Δ	Δ
minisono	minisono L3-12/C1-6	Δ	O	Δ	X	O	O	O
Others	US-Guider	Δ	O	Δ	X	X	O	O
	Strain relief&Cable	Δ	O	Δ	X	O	O	X

※ CAUTION ※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Disinfectant						
		AIDAL PLUS	Ster-Bac Blu	Hibitane	Metricide 14	Sporicidin	Omnicide -FG2	Nuclear
Manufacturer		Whiteley Medical	Ecolab	COOPERS	Metrex	CONTEC	Interhatch	Ecolab
Level		HLD	-	-	HLD	HLD	-	-
Active Ingredients		Glutaraldehyde	Quaternary Ammonium	Clorhexidine gluconate	Glutaraldehyde	Phenol	Glutaraldehyde	Glutaraldehyde
FDA Approval		Y	N	N	Y	Y	Y	N
Disinfectant Type		Liquid	Liquid	Liquid	Liquid	Liquid	Liquid	Liquid
Convex	C1-6/C1-6T/C1-6i/SC1-6 /SC1-4HS/SC1-4H/ SC1-6H/ SC2-9H/SC1-7H	O	Δ	Δ	Δ	X	Δ	Δ
	C1-6CT	O	Δ	Δ	Δ	O	Δ	Δ
Micro-Convex	C5-8	Δ	O	O	O	Δ	Δ	Δ
	C5-8N/C5-8NT/SC2-11H	O	Δ	Δ	Δ	X	Δ	Δ
Linear	L3-8/L3-8i/L3-12/L3-12T /L3-12i/L3-12H/ L3-12H ^{WD} /L3-2X/L8-17 /L8-17i/L8-17X/L3-15H	O	Δ	Δ	Δ	X	Δ	Δ
	L10-25H/SL3-19H /SL3-19X	O	Δ	Δ	Δ	Δ	Δ	Δ
Phased	MP1-5X/ SP1-5/SP1-5i/ SP1-5X/P1-5CT/SP1-5T/ SP3-8/SP3-8T	O	Δ	Δ	Δ	X	Δ	Δ
Volume	VC1-6/VC1-6T/SVC1-6/ SVC1-6H/VE3-10H/ SVC1-8H	O	Δ	Δ	Δ	Δ	Δ	Δ
Endo	EV3-10/EV3-10T/ EV3-10H/EV3-10X	O	Δ	Δ	Δ	X	Δ	Δ
	EC3-10/EC3-10T/ EC3-10H/EC3-10X/ EC(V)2-11H	O	Δ	Δ	Δ	X	Δ	Δ
	EN3-10	Δ	Δ	Δ	Δ	Δ	O	O
	E3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	E3-10H	Δ	Δ	Δ	Δ	Δ	Δ	Δ
IO	IO3-12/IO8-17/IO8-17T	O	Δ	Δ	Δ	X	Δ	Δ
CW	CW2.0/CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	Δ	Δ	Δ	Δ	Δ	Δ	Δ
minisono	minisono L3-12/C1-6	O	Δ	Δ	Δ	X	Δ	Δ
Others	US-Guider	O	Δ	Δ	Δ	O	Δ	Δ
	Strain relief&Cable	X	Δ	Δ	Δ	X	Δ	Δ

※ CAUTION ※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Disinfectant						
		Sani-Cloth HB	PeraSafe	Super Sani-Cloth	Cleanisept wipes	CLEANISEPT® WIPES forte	Protex	Tristel Duo
Manufacturer		NICE PAK	VWR	NICE PAK	Dr. Schumacher GmbH	Dr. Schumacher GmbH	PARKER	Tristel Solutions Ltd.
Level		LLD	HLD	-	LLD	LLD	LLD	HLD
Active Ingredients		Quaternary Ammonium	Peracetic	Alcohol	Quaternary Ammonium	Quaternary Ammonium	Quaternary Ammonium	Chlorine Dioxide
FDA Approval		N	N	N	N	N	Y	N
Disinfectant Type		Wipes	Liquid	Wipes	Wipes	Wipes	Spray	Spray
Convex	C1-6/C1-6T/C1-6i/SC1-6/SC1-4HS/SC1-4H/SC1-6H/ SC2-9H/SC1-7H	Δ	Δ	O	O	O	O	O
	C1-6CT	Δ	Δ	Δ	Δ	O	Δ	O
Micro-Convex	C5-8	O	O	Δ	Δ	Δ	Δ	Δ
	C5-8N/C5-8NT/SC2-11H	Δ	Δ	O	O	O	O	O
Linear	L3-8/L3-8i/L3-12/L3-12T/L3-12i/L3-12H/L3-12H ^{WD} /L3-2X/L8-17/L8-17i/L8-17X/L3-15H	Δ	Δ	O	O	O	O	O
	L10-25H/SL3-19H/SL3-19X	Δ	Δ	O	O	O	Δ	O
Phased	MP1-5X/ SP1-5/SP1-5i/SP1-5X/P1-5CT/SP1-5T/SP3-8/SP3-8T	Δ	Δ	O	O	O	O	O
Volume	VC1-6/VC1-6T/SVC1-6/SVC1-6H/VE3-10H/SVC1-8H	Δ	Δ	O	O	O	Δ	O
Endo	EV3-10/EV3-10T/EV3-10H/EV3-10X	Δ	Δ	O	O	O	O	O
	EC3-10/EC3-10T/EC3-10H/EC3-10X/EC(V)2-11H	Δ	Δ	O	O	O	O	O
	EN3-10	Δ	Δ	O	Δ	Δ	Δ	Δ
	E3-10	Δ	Δ	O	O	O	Δ	Δ
	E3-10H	Δ	Δ	O	O	O	Δ	Δ
IO	IO3-12/IO8-17/IO8-17T	Δ	Δ	O	O	O	O	O
CW	CW2.0/CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	Δ	O	Δ	Δ	Δ	Δ	Δ
minisono	minisono L3-12/C1-6	Δ	Δ	O	O	O	O	O
Others	US-Guider	Δ	Δ	O	O	O	O	X
	Strain relief&Cable	Δ	Δ	O	O	O	X	O

※ CAUTION ※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Disinfectant						
		SteriMax® Sporicide Wipes	SteriMax® Wipes	SteriMax Sporicide	UltraSan Broad Spectrum	CaviCide	CaviWipes	Metricide OPA Plus
Manufacturer		Aseptix	Aseptix	Aseptix	Aseptix	Metrex	Metrex	Metrex
Level		-	-	-	-	-	LLD,ILD	HLD
Active Ingredients		Hydrogen peroxide	Hydrogen peroxide	Hydrogen peroxide	Hydrogen peroxide	Isopropanol	Isopropanol	Orthophthalalde hyde
FDA Approval		N	N	N	N	Y	Y	Y
Disinfectant Type		Wipes	Wipes	Spray	Spray	Spray	Wipes	Liquid
Convex	C1-6/C1-6T/C1-6i/SC1-6 /SC1-4HS/SC1-4H/ SC1-6H/ SC2-9H/SC1-7H	O	X	X	X	X	O	O
	C1-6CT	O	O	Δ	Δ	O	O	Δ
Micro- Convex	C5-8	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	C5-8N/C5-8NT/SC2-11H	O	X	X	X	X	O	O
Linear	L3-8/L3-8i/L3-12/L3-12T /L3-12i/L3-12H/ L3-12H ^{WD} /L3-2X/L8-17 /L8-17i/L8-17X/L3-15H	O	X	X	X	X	O	O
	L10-25H/SL3-19H /SL3-19X	Δ	Δ	Δ	Δ	Δ	Δ	O
Phased	MP1-5X/ SP1-5/SP1-5i/ SP1-5X/P1-5CT/SP1-5T/ SP3-8/SP3-8T	O	X	X	X	X	O	O
Volume	VC1-6/VC1-6T/SVC1-6/ SVC1-6H/VE3-10H/ SVC1-8H	Δ	Δ	Δ	Δ	Δ	Δ	O
Endo	EV3-10/EV3-10T/ EV3-10H/EV3-10X	O	X	X	X	X	O	O
	EC3-10/EC3-10T/ EC3-10H/EC3-10X/ EC(V)2-11H	O	X	X	X	X	O	O
	EN3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	E3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	E3-10H	Δ	Δ	Δ	Δ	Δ	Δ	Δ
IO	IO3-12/IO8-17/IO8-17T	O	X	X	X	X	O	O
CW	CW2.0/CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	Δ	Δ	Δ	Δ	Δ	Δ	Δ
minisono	minisono L3-12/C1-6	O	X	X	X	X	O	O
Others	US-Guider	X	X	X	X	X	O	O
	Strain relief&Cable	O	O	X	X	O	O	O

※ CAUTION ※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Disinfectant						
		Metricide	Meliseptol Wipes sensitive	Meliseptol Wipes ultra	Revital-OX Resert	Meliseptol Foam pure	Sporox II	Protex wipes
Manufacturer		Metrex	B. Braun	B. Braun	STERIS	B. Braun	SultanHealthcare	PARKER
Level		HLD	-	-	HLD	-	HLD	-
Active Ingredients		Glutaraldehyde	Propan-1-ol	Quaternary Ammonium	Hydrogen Peroxide	Propan-1-ol	Hydrogen Peroxide	Ammonium chloride
FDA Approval		Y	N	N	Y	N	Y	Y
Disinfectant Type		Liquid	Wipes	Wipes	Liquid	Liquid	Liquid	Wipes
Convex	C1-6/C1-6T/C1-6i/SC1-6 /SC1-4HS/SC1-4H/ SC1-6H/ SC2-9H/SC1-7H	O	X	O	O	O	O	O
	C1-6CT	Δ	Δ	Δ	O	O	X	O
Micro-Convex	C5-8	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	C5-8N/C5-8NT/SC2-11H	O	X	O	O	O	O	O
Linear	L3-8/L3-8i/L3-12/L3-12T /L3-12i/L3-12H/ L3-12H ^{WD} /L3-2X/L8-17 /L8-17i/L8-17X/L3-15H	O	X	O	O	O	O	O
	L10-25H/SL3-19H /SL3-19X	Δ	O	O	O	O	O	O
Phased	MP1-5X/ SP1-5/SP1-5i/ SP1-5X/P1-5CT/SP1-5T/ SP3-8/SP3-8T	O	X	O	O	O	O	O
Volume	VC1-6/VC1-6T/SVC1-6/ SVC1-6H/VE3-10H/ SVC1-8H	Δ	O	O	O	O	O	O
Endo	EV3-10/EV3-10T/ EV3-10H/EV3-10X	O	X	O	O	O	O	O
	EC3-10/EC3-10T/ EC3-10H/EC3-10X/ EC(V)2-11H	O	X	O	O	O	O	O
	EN3-10	Δ	Δ	Δ	Δ	Δ	X	Δ
	E3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	E3-10H	Δ	Δ	Δ	Δ	Δ	Δ	Δ
IO	IO3-12/IO8-17/IO8-17T	O	X	O	O	O	O	O
CW	CW2.0/CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	Δ	Δ	Δ	Δ	Δ	Δ	Δ
minisono	minisono L3-12/C1-6	O	X	O	O	O	O	O
Others	US-Guider	X	Δ	Δ	Δ	Δ	Δ	Δ
	Strain relief&Cable	O	O	O	X	O	Δ	Δ

※ CAUTION ※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Disinfectant					
		Sani-cloth AF3	Acrylan	Descogen Liquid	Cutasept F	Cutasept G	Mikrobac Virucidal Tissues
Manufacturer		NICE PAK	ANTISEPTICA	ANTISEPTICA	BODE Chemie GmbH	BODE Chemie GmbH	BODE Chemie GmbH
Level		LLD,ILD	-	-	-	-	-
Active Ingredients		Quaternary Ammonium	Quaternary ammonium compounds	Potassium peroxymonosulfate	Propan-2-ol	Propan-2-ol	ammonium chloride
FDA Approval		Y	N	N	N	N	N
Disinfectant Type		Wipes	Liquid	Liquid	Liquid	Liquid	Wipes
Convex	C1-6/C1-6T/C1-6i/SC1-6/SC1-4HS/SC1-4H/SC1-6H/ SC2-9H/SC1-7H	O	X	O	X	X	O
	C1-6CT	O	Δ	Δ	X	X	O
Micro-Convex	C5-8	Δ	Δ	Δ	Δ	Δ	Δ
	C5-8N/C5-8NT/SC2-11H	O	X	O	X	X	O
Linear	L3-8/L3-8i/L3-12/L3-12T/L3-12i/L3-12H/L3-12H ^{WD} /L3-2X/L8-17/L8-17i/L8-17X/L3-15H	O	X	O	X	X	O
	L10-25H/SL3-19H/SL3-19X	O	O	O	O	Δ	O
Phased	MP1-5X/ SP1-5/SP1-5i/SP1-5X/P1-5CT/SP1-5T/SP3-8/SP3-8T	O	X	O	X	X	O
Volume	VC1-6/VC1-6T/SVC1-6/SVC1-6H/VE3-10H/SVC1-8H	O	O	O	O	Δ	O
Endo	EV3-10/EV3-10T/EV3-10H/EV3-10X	O	X	O	X	X	O
	EC3-10/EC3-10T/EC3-10H/EC3-10X/EC(V)2-11H	O	X	O	X	X	O
	EN3-10	Δ	Δ	Δ	Δ	Δ	Δ
	E3-10	Δ	Δ	Δ	Δ	Δ	Δ
	E3-10H	Δ	Δ	Δ	Δ	Δ	Δ
IO	IO3-12/IO8-17/IO8-17T	O	X	O	X	X	O
CW	CW2.0/CW5.0	O	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	Δ	Δ	Δ	Δ	Δ	Δ
minisono	minisono L3-12/C1-6	O	X	O	X	X	O
Others	US-Guider	Δ	Δ	Δ	Δ	Δ	Δ
	Strain relief&Cable	Δ	Δ	Δ	Δ	Δ	Δ

※CAUTION※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Disinfectant			Pre-Cleaner	
		SONO Ultrasound and Disinfecting Wipe	Bacillol 30 Sensitive Foam	Bacillol 30 Sensitive Tissues	T-spray	T-spray II
Manufacturer		ADVANCED ULTRASOUND SOLUTIONS INC.	BODE Chemie GmbH	BODE Chemie GmbH	Pharmaceutical Innovations, INC	Pharmaceutical Innovations, INC
Level		-	-	-	-	-
Active Ingredients		ammonium chloride	Ethanol Propan-2-ol	Ethanol Propan-2-ol	Alkyl dimethyl benzyl ammonium chloride	Alkyl dimethyl benzyl ammonium chloride
FDA Approval		N	N	N	Y	Y
Disinfectant Type		Wipes	Liquid	Wipes	Spray	Spray
Convex	C1-6/C1-6T/C1-6i/SC1-6 /SC1-4HS/SC1-4H/ SC1-6H/ SC2-9H/SC1-7H	X	O	O	Δ	Δ
	C1-6CT	O	O	O	Δ	Δ
Micro-Convex	C5-8	Δ	Δ	Δ	O	O
	C5-8N/C5-8NT/SC2-11H	X	O	O	Δ	Δ
Linear	L3-8/L3-8i/L3-12/L3-12T /L3-12i/L3-12H/ L3-12H ^{WD} /L3-2X/L8-17 /L8-17i/L8-17X/L3-15H	X	O	O	Δ	Δ
	L10-25H/SL3-19H /SL3-19X	O	O	O	Δ	O
Phased	MP1-5X/ SP1-5/SP1-5i/ SP1-5X/P1-5CT/SP1-5T/ SP3-8/SP3-8T	X	O	O	Δ	Δ
Volume	VC1-6/VC1-6T/SVC1-6/ SVC1-6H/VE3-10H/ SVC1-8H	O	O	O	Δ	O
Endo	EV3-10/EV3-10T/ EV3-10H/EV3-10X	X	O	O	Δ	Δ
	EC3-10/EC3-10T/ EC3-10H/EC3-10X/ EC(V)2-11H	X	O	O	Δ	Δ
	EN3-10	Δ	Δ	Δ	O	Δ
	E3-10	Δ	Δ	Δ	O	Δ
	E3-10H	Δ	Δ	Δ	O	Δ
IO	IO3-12/IO8-17/IO8-17T	X	O	O	Δ	Δ
CW	CW2.0/CW5.0	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	Δ	Δ	Δ	Δ	Δ
minisono	minisono L3-12/C1-6	X	O	O	Δ	Δ
Others	US-Guider	Δ	Δ	Δ	Δ	Δ
	Strain relief&Cable	Δ	Δ	Δ	Δ	Δ

※ CAUTION ※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Pre-Cleaner					
		Enzol	Theracide Plus(S5N)	Transeptic	Metrizyme	Klenzyme	McKesson
Manufacturer		Johnson and Johnson	Burnishine Products	PARKER	Metrex	STERIS	McKesson
Level		-	-	-	-	-	-
Active Ingredients		Subtilisins(proteolytic enzymes)	Isopropyl alcohol	Isopropyl alcohol	Propylene glycol	Enzymes	PCMX (Chloroxylenol)
FDA Approval		Y	N	Y	Y	Y	Y
Disinfectant Type		Liquid	Liquid	Liquid	Liquid	Liquid	Liquid
Convex	C1-6/C1-6T/C1-6i/SC1-6/SC1-4HS/SC1-4H/SC1-6H/ SC2-9H/SC1-7H	O	Δ	Δ	O	Δ	Δ
	C1-6CT	Δ	Δ	Δ	O	Δ	Δ
Micro-Convex	C5-8	Δ	Δ	Δ	Δ	Δ	Δ
	C5-8N/C5-8NT/SC2-11H	O	Δ	Δ	O	Δ	Δ
Linear	L3-8/L3-8i/L3-12/L3-12T/L3-12i/L3-12H/L3-12H ^{WD} /L3-2X/L8-17/L8-17i/L8-17X/L3-15H	O	Δ	Δ	O	Δ	Δ
	L10-25H/SL3-19H/SL3-19X	O	O	O	O	Δ	Δ
Phased	MP1-5X/ SP1-5/SP1-5i/SP1-5X/P1-5CT/SP1-5T/SP3-8/SP3-8T	O	Δ	Δ	O	Δ	Δ
Volume	VC1-6/VC1-6T/SVC1-6/SVC1-6H/VE3-10H/SVC1-8H	O	O	O	O	Δ	Δ
Endo	EV3-10/EV3-10T/EV3-10H/EV3-10X	O	Δ	Δ	O	Δ	Δ
	EC3-10/EC3-10T/EC3-10H/EC3-10X/EC(V)2-11H	O	Δ	Δ	O	Δ	Δ
	EN3-10	Δ	Δ	Δ	O	O	O
	E3-10	Δ	Δ	Δ	Δ	Δ	Δ
	E3-10H	Δ	Δ	Δ	Δ	Δ	Δ
IO	IO3-12/IO8-17/IO8-17T	O	Δ	Δ	O	Δ	Δ
CW	CW2.0/CW5.0	Δ	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	O	Δ	Δ	Δ	Δ	Δ
minisono	minisono L3-12/C1-6	O	Δ	Δ	O	Δ	Δ
Others	US-Guider	O	Δ	O	X	Δ	Δ
	Strain relief&Cable	O	Δ	O	O	Δ	Δ

※ CAUTION ※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Pre-Cleaner				Gel	
		Cidezyme	Endozyme	EmPower	Ultrasound probe cleaning wipes	Aquasonic100	ULTRASOUND GEL
Manufacturer		Johnson and Johnson	RUHOF Corporation	Metrex	Nanosonics	PARKER	Ultragel Hungary 2000 Ltd.
Level		-	-	-	-	-	-
Active Ingredients		Proteolytic Enzymes	Propanol	Propylene glycol	PHENOXYETHANOL	-	-
FDA Approval		Y	Y	Y	N	N	N
Disinfectant Type		Liquid	Liquid	Liquid	Wipes	Gel	Gel
Convex	C1-6/C1-6T/C1-6i/SC1-6/SC1-4HS/SC1-4H/SC1-6H/ SC2-9H/SC1-7H	Δ	Δ	O	O	O	O
	C1-6CT	Δ	Δ	O	Δ	O	Δ
Micro-Convex	C5-8	Δ	Δ	Δ	Δ	O	O
	C5-8N/C5-8NT/SC2-11H	Δ	Δ	O	O	O	O
Linear	L3-8/L3-8i/L3-12/L3-12T/L3-12i/L3-12H/L3-12H ^{WD} /L3-2X/L8-17/L8-17i/L8-17X/L3-15H	Δ	Δ	O	O	O	O
	L10-25H/SL3-19H/SL3-19X	Δ	O	O	O	O	Δ
Phased	MP1-5X/ SP1-5/SP1-5i/SP1-5X/P1-5CT/SP1-5T/SP3-8/SP3-8T	Δ	Δ	O	O	O	O
Volume	VC1-6/VC1-6T/SVC1-6/SVC1-6H/VE3-10H/SVC1-8H	Δ	O	O	O	O	Δ
Endo	EV3-10/EV3-10T/EV3-10H/EV3-10X	Δ	Δ	O	O	O	O
	EC3-10/EC3-10T/EC3-10H/EC3-10X/EC(V)2-11H	Δ	Δ	O	O	O	O
	EN3-10	O	Δ	Δ	Δ	Δ	Δ
	E3-10	Δ	Δ	Δ	Δ	O	O
	E3-10H	Δ	Δ	Δ	Δ	O	O
IO	IO3-12/IO8-17/IO8-17T	Δ	Δ	O	O	O	O
CW	CW2.0/CW5.0	Δ	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	O	Δ	Δ	Δ	O	Δ
minisono	minisono L3-12/C1-6	Δ	Δ	O	O	O	O
Others	US-Guider	Δ	Δ	O	Δ	O	Δ
	Strain relief&Cable	Δ	Δ	O	Δ	O	Δ

※CAUTION※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

***NOTE**

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23

Transducer Disinfectant Material Compatibility Table

[Rev.23]

HLD : High level disinfectant / ILD : Intermediate level disinfectant / LLD : Low level disinfectant

O : Approved / X : Not Approved / Δ : Not Tested

Product Type		Gel				Trophon® EPR	Antigermix ^{Note)}	V-PRO ^{Note)}
		Ultrasound Transmission gel	Scan	Sonogel	Super sonic			
Manufacturer		Ecosonic	PARKER	SONOGEL Vertriebs GmbH	Sung Heung Corporation			
Level		-	-	-	-			
Active Ingredients		-	-	-	-			
FDA Approval		N	N	N	N			
Disinfectant Type		Gel	Gel	Gel	Gel	Device	Device	Device
Convex	C1-6/C1-6T/C1-6i/SC1-6 /SC1-4HS/SC1-4H/ SC1-6H/ SC2-9H/SC1-7H	O	Δ	O	O	O	O	O
	C1-6CT	Δ	Δ	Δ	Δ	Δ	Δ	O
Micro- Convex	C5-8	Δ	O	O	Δ	Δ	Δ	Δ
	C5-8N/C5-8NT/SC2-11H	O	Δ	O	O	O	O	O
Linear	L3-8/L3-8i/L3-12/L3-12T /L3-12i/L3-12H/ L3-12H ^{WD} /L3-2X/L8-17 /L8-17i/L8-17X/L3-15H	O	Δ	O	O	O	O	O
	L10-25H/SL3-19H /SL3-19X	Δ	Δ	O	Δ	O	X	O
Phased	MP1-5X/ SP1-5/SP1-5i/ SP1-5X/P1-5CT/SP1-5T/ SP3-8/SP3-8T	O	Δ	O	O	O	O	O
Volume	VC1-6/VC1-6T/SVC1-6/ SVC1-6H/VE3-10H/ SVC1-8H	Δ	Δ	O	Δ	O	X	O
Endo	EV3-10/EV3-10T/ EV3-10H/EV3-10X	O	Δ	O	O	X	O	O
	EC3-10/EC3-10T/ EC3-10H/EC3-10X/ EC(V)2-11H	O	Δ	O	O	O	O	O
	EN3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	E3-10	Δ	Δ	Δ	Δ	Δ	Δ	Δ
	E3-10H	Δ	Δ	Δ	Δ	Δ	Δ	Δ
IO	IO3-12/IO8-17/IO8-17T	O	Δ	O	O	O	O	O
CW	CW2.0/CW5.0	Δ	Δ	Δ	Δ	Δ	Δ	Δ
TEE	TEE3-7	Δ	Δ	Δ	Δ	Δ	Δ	Δ
minisono	minisono L3-12/C1-6	O	Δ	O	O	X	O	X
Others	US-Guider	Δ	Δ	Δ	X	Δ	Δ	Δ
	Strain relief&Cable	Δ	Δ	O	O	Δ	X	O

※ CAUTION ※

- Only wipe the cable, strain relief and parts of the transducer. Do not wipe the connector of the transducer.
- Do not immerse transducers in a solution for longer than one hour, unless under disinfectant manufacturer's instructions.
- Using a non-recommended disinfectant or not following the recommended disinfection method can damage and/or discolor the transducer and will void the transducer warranty.

*NOTE

- Antigermix, V-PRO : Slight discoloration of the housing and cable can occur. These changes have no impact on device performance

Alpinion Medical Systems Co., Ltd.

e-mail : marketing@alpinion.com TEL : +82-2-3282-0900

Copyright© 2017-2022 ALPINION MEDICAL SYSTEMS CO., LTD. All rights reserved

MKT-EL-TS-LF-E2112-V23